



Raja Peary Mohan College

NAAC accredited with B grade (2.7); ISO 9001:2015

Uttarpara, Hooghly, West Bengal Pin- 712 258

Ph (033)26630881, Fax(033) 26634155

Website: www.rpmcollege.org , e-mail: rajapearymohancollege@gmail.com

Dated: December 14, 2019

Notice Inviting Tender No. : RPMC/Tender/82/2019-20

Sealed tender invited on behalf of the Raja Peary Mohan College from the Original Equipment Manufacturer or their authorized dealers for the **Supply & Installation of Lab Equipments** for the laboratory of the institute as per the specifications in. The tender document can be downloaded from the Institute website at URL Link: <http://www.rpmcollege.org>.

1. Schedule

Date of Issue/Publishing	December 14, 2019
Last Date and Time for submission of Bids	December 23, 2019 03:00 PM
Cover heading	The cover containing the quotation should be subscribed as Bid for Supply & Installation of Lab Equipments.
Bid Validity days	90 days (From last date of opening of tender)
Item	<u>SPECIFICATIONS FOR PIPETTE</u> <ol style="list-style-type: none">1. The one handed volume adjustment to avoid unintentional adjustment2. IvD-conformity3. Fully autoclavable4. Light weight5. Must be capable to regulate the piston movement speed to pipette higher viscous (i.e glycerol) and lower viscous (i.e chloroform) samples6. Calibration seal for optimal GLP specifications7. Chemical resistant8. Must have spring loaded tip cone9. Piston should be made of organic polymer and resistant to acids, bases, dyes, sunlight etc10. 4-digit magnifying display11. Multichannel pipettes must have facility to adjust its channel according to the wells of plate.12. Single-channel adjustable variable volume pipette13. Should have NABL accredited lab facility for providing calibration service on demand14. Should provide pipette stand for 6 pipettes15. Volume range required: 0.5-10 μL, 10-100 μL, 100-1000 μL

2. Two Bid System :

- i. Technical bid consisting of all technical details alongwith commercial terms and conditions; and
- ii. Financial bid indicating item-wise price for the items mentioned in the technical bid.
- iii. Technical bid and financial bid should be sealed in separate covers duly superscribed and both these sealed covers are to be put in a bigger cover which should also be sealed and duly superscribed as **"Technical Bid for the supply of installation of Lab Equipments" due on December 23, 2019, 03:00 PM**. Technical bids shall be opened at the first instance and evaluated by technical committee. At the second stage financial bids of the only technically qualified bidders shall be opened for financial evaluation and ranking before awarding the contract. Mixing price bid with technical bid will disqualify your bid for further evaluation.

3. Submission of tender:

Offers addressed to the 'Principal, Raja Peary Mohan College' and valid for 90 days (From last date of opening of tender) should reach the office on or before the last date and time. Tenders received late shall not be considered.


PRINCIPAL
Raja Peary Mohan College
Uttarakhand, India